

98.9% overall concordance

Routine reference methods, including further analysis by ddPCR, Idylla™ retest, NGS

	Codon 12	Codon 13	Codon 59	Codon 61	Codon 117	Codon 146	No mutations	Total
Idylla™ KRAS Mutation Assay	<b>142</b>							142
		<b>38</b>						38
			<b>5</b>					5
	<b>1</b>			<b>15</b>			<b>1</b>	17
					<b>4</b>			4
								20
						<b>20</b>		19
	<b>1</b>	<b>2</b>					<b>133</b>	136
TOTAL	144	40	5	15	4	20	134	362

### Idylla™ KRAS Posters & Publications

- Maertens G. et al. A solution for same-day extended RAS testing. Poster ESMO 2015
- Vandembroucke I. et al. A rapid and fully automated multiplex assay for KRAS-BRAF mutations with high mutation sensitivity using novel selective amplification and detection technologies. Poster AACR 2014
- Solassol J. et al. Multi-Center Evaluation of the Fully Automated PCR-Based Idylla™ KRAS Mutation Assay for Rapid KRAS Mutation Status Determination on Formalin-Fixed Paraffin-Embedded Tissue of Human Colorectal Cancer. PLOS ONE 2016

### Research applications

- Dario de Biase. et al. Fully automated PCR detection of KRAS mutations on pancreatic endoscopic ultrasound fine-needle aspirates. J Clin Path 2016.

### Disclaimer

Biocartis trademark and logo are trademarks belonging to Biocartis and are used and registered in Europe. Idylla is a registered trademark in the United States (US) and other countries. Idylla trademark and logo are used trademarks belonging to Biocartis. Idylla™ platform and Idylla™ KRAS Mutation Test are CE-marked IVDs in Europe. Idylla™ KRAS Mutation Assay is available for Research Use Only, not for use in diagnostic procedures. Idylla™ is not for sale in USA, Canada and some other countries. Please check availability with the local Biocartis sales representative.

The **Idylla™ KRAS Mutation Test**, performed on the Biocartis Idylla™ system, is an *in vitro* diagnostic Test for the qualitative detection of 21 mutations in **codons 12, 13, 59, 61, 117 and 146** of the *KRAS* gene. The Idylla™ KRAS Mutation Test, **from sample-to-result**, starts with formalin-fixed, paraffin-embedded (FFPE) human tissue from metastatic colorectal cancers to liberate DNA for subsequent real-time PCR amplification and detection.



### Features

KRAS mutation detection		
	G12C	(c.34G>T)
	G12R	(c.34G>C)
Codon 12 (exon 2)	G12S	(c.34G>A)
	G12A	(c.35G>C)
	G12D	(c.35G>A)
	G12V	(c.35G>T)
Codon 13 (exon 2)	G13D	(c.38G>A)
	A59E	(c.176C>A)
Codon 59 (exon 3)	A59G	(c.176C>G)
	A59T	(c.175G>A)
	Q61K	(c.181C>A; c.180_181delinsAA)
Codon 61 (exon 3)	Q61L	(c.182A>T)
	Q61R	(c.182A>G)
	Q61H	(c.183A>C; c.183A>T)
Codon 117 (exon 4)	K117N	(c.351A>C; c.351A>T)
	A146P	(c.436G>C)
Codon 146 (exon 4)	A146T	(c.436G>A)
	A146V	(c.437C>T)
	KRAS Total (acting as Sample Processing Control)	
Specimen requirements		
Sample Type	FFPE tissue sections (5 to 10µm)	
Neoplastic cells	≥10%, if less macrodissection is required	
Tissue area	50-600mm² (5µm) 25-300mm² (10µm)	
Performance		
Analytical Sensitivity	LOD ≤5% for all KRAS mutations	

Between Laboratory Reproducibility (480 results at 3 sites)	100% agreement for 5% KRAS G12D 100% agreement for 5% KRAS G12S 100% agreement for 5% KRAS G12V 100% agreement for 50% KRAS G13D
Between Lot Reproducibility (375 results on 3 lots)	100% agreement for 5% KRAS G12A 100% agreement for 5% KRAS G12D 100% agreement for 5% KRAS G12S 100% agreement for 5% KRAS G12V 100% agreement for 5% KRAS G13D
Total turnaround time	
Time	120 minutes

**Accuracy - Clinical Performance Evaluation**

96.7% overall percent agreement was obtained during the clinical performance evaluation comparing Idylla™ with a reference method based on RT-PCR.

96.7% overall concordance

		Reference test								
		G12A	G12C	G12D	G12R	G12S	G12V	G13D	No mutation detected	Total
Idylla™ KRAS Mutation Test	G12A	<b>6</b>								6
	G12C		<b>6</b>							6
	G12D			<b>25</b>						25
	G12R				<b>3</b>				<b>1</b>	4
	G12S					<b>6</b>				6
	G12V					<b>1</b>	<b>15</b>		<b>1</b>	17
	G13D					<b>1</b>		<b>16</b>	<b>3</b>	20
	No mutation detected	<b>1</b>							<b>97</b>	98
	A59E/G/T									1
	Q61H/H2									3
Q61K/K2									0	
Q61L/R									0	
K117N/N2		1							2	3
A146T/V/P							1		4	5
Totals		7	7	25	3	8	16	16	112	194

Note: the Reference test is not designed to pick up mutations in codon 59, 61, 117 and 146

\*Due to limitations in available material (n=2) or insufficient DNA quality (n=3), only one of the 6 discordant results could be resolved; NGS confirmed the G12V result called by Idylla™

↓  
Discordant analysis by NGS

100% overall concordance

		Reference test and further analysis by NGS								
		G12A	G12C	G12D	G12R	G12S	G12V	G13D	No mutation detected	Total
Idylla™ KRAS Mutation Test	G12A	<b>6</b>								6
	G12C		<b>6</b>							6
	G12D			<b>25</b>						25
	G12R				<b>3</b>					3
	G12S					<b>6</b>				6
	G12V					<b>1</b>	<b>16</b>			17
	G13D					<b>1</b>		<b>16</b>		17
	No mutation detected								<b>97</b>	97
	A59E/G/T									1
	Q61H/H2									3
Q61K/K2									0	
Q61L/R									0	
K117N/N2			1						2	3
A146T/V/P							1		4	5
Totals		6	7	25	3	8	17	16	107	189

**Multi-center evaluation of the fully-automated PCR-based Idylla™ KRAS Mutation Assay for rapid KRAS mutation status determination on formalin-fixed paraffin-embedded tissue of human colorectal cancer.**

Solassol J. et al. PLOS ONE 2016.

95.9% overall concordance

		Routine reference methods*							
		Codon 12	Codon 13	Codon 59	Codon 61	Codon 117	Codon 146	No mutation detected	Total
Idylla™ KRAS Mutation Assay	Codon 12	<b>138</b>						<b>3</b>	141
	Codon 13		<b>38</b>						38
	Codon 59			<b>5</b>					5
	Codon 61	<b>1</b>			<b>14</b>			<b>5</b>	20
	Codon 117					<b>4</b>			4
	Codon 146						<b>18</b>	<b>1</b>	19
No mutation	<b>3</b>	<b>2</b>				<b>1</b>	<b>129</b>	135	
TOTAL	142	40	5	14	4	19	138	362	

Mutations not picked up by Idylla™ or the reference method (n=8) are not included in the % agreement calculation. Idylla does not pick up the following low prevalent (<1%) mutations: 4xG13C, 1xG13R, 1xG12F (samples excluded from table). Roche Cobas does not pick up mutations in codon 146 (2x). \*Different reference methods were used: cobas® KRAS Mutation Test (Roche), Ion Torrent AmpliSeq™ Colon and Lung Cancer Research Panel (Life Technologies), theascreen® KRAS Pyro® Kit (Qiagen), theascreen® RAS Extension Pyro Kit (Qiagen), HRM screening and pyrosequencing, Sanger sequencing, HRM screening and Sanger sequencing; for the analysis, when Idylla™ identified a specific mutation in codon 12, 13 or 61, and the cobas® KRAS Mutation Test (Roche) reported a "codon 12/13" or "codon 61" result, both results were considered identical.

↓  
Discordant analysis by ddPCR, Idylla™ retest and NGS